

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.669 °C                  | 11.0 °C            | 0.331 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.065 °C                  | 19.8 °C            | -0.265 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 29.310 °C                  | 29.0 °C            | -0.310 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 67 %HR                     | 67.656 %HR                 | 71.54 %HR          | 3.884 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 88 %HR                     | 90.151 %HR                 | 85.97 %HR          | -4.181 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |