

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	23.2 μm	-0.1 μm	± 1.2 μm	±	5.8E-01 μm
51.6 μm	52.0 μm	0.4 μm	± 2.5 μm	±	5.8E-01 μm
124.6 μm	124.8 μm	0.2 μm	± 3.2 μm	±	5.8E-01 μm
250.3 μm	251.0 μm	0.7 μm	± 4.5 μm	±	5.8E-01 μm
486.0 μm	486.7 μm	0.7 μm	± 6.9 μm	±	5.8E-01 μm
963.0 μm	962.1 μm	-0.9 μm	± 12 μm	±	6.0E-01 μm