

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04955	0.450	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.09980	0.200	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14992	0.080	± 0.6	± 4.3E-01 mm					
0 a 1	0.99956	0.440	± 1	± 4.3E-01 mm					
0 a 2	1.99972	0.280	± 1.4	± 4.8E-01 mm					
0 a 3	2.99905	0.950	± 1.8	± 4.8E-01 mm					
0 a 4	3.99925	0.750	± 2.2	± 4.8E-01 mm					
0 a 5	4.99911	0.890	± 2.6	± 4.8E-01 mm					
0 a 6	5.99974	0.260	± 3	± 3.6E-01 mm					
0 a 7	6.99916	0.840	± 3.4	± 3.6E-01 mm					
0 a 8	7.99944	0.560	± 3.8	± 3.6E-01 mm					
0 a 9	8.99972	0.280	± 4.2	± 3.6E-01 mm					
0 a 10	10.00015	-0.150	± 4.6	± 3.6E-01 mm					
0 a 11	10.99897	1.030	± 5	± 3.6E-01 mm					
0 a 12	11.99912	0.880	± 5.4	± 3.6E-01 mm					
0 a 13	12.99941	0.590	± 5.8	± 3.6E-01 mm					
0 a 14	13.99958	0.420	± 6.2	± 3.6E-01 mm					
0 a 15	14.99948	0.520	± 6.6	± 3.6E-01 mm					