

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.221 °C	0.1 °C	0.321 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.623 °C	15.5 °C	-0.123 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.3 %HR	-0.270 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.222 %HR	48.9 %HR	-1.322 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.5 %HR	-0.6 %HR	± 2.5 %HR	± 1.9E+00 %HR