

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
10 °C	10.007 °C	10.5 °C	0.493 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	29.5 °C	-0.963 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.118 °C	49.5 °C	-0.618 °C	± 1.0 °C	± 1.1E+00 °C
30 %HR	30.570 %HR	26.0 %HR	-4.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	44.0 %HR	-6.049 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	69.0 %HR	-6.560 %HR	± 5.0 %HR	± 1.9E+00 %HR