

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	11.0 °C	0.993 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.0 °C	-0.463 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.118 °C	49.0 °C	-1.118 °C	± 1.0 °C	± 1.1E+00 °C
30 %HR	30.570 %HR	26.0 %HR	-4.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	47.0 %HR	-3.049 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	73.0 %HR	-2.560 %HR	± 5.0 %HR	± 1.9E+00 %HR