

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	9.0 °C	-1.007 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.463 °C	29.0 °C	-1.463 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.118 °C	48.0 °C	-2.118 °C	± 1.0 °C	± 1.1E+00 °C
30 %HR	30.570 %HR	25.0 %HR	-5.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	40.0 %HR	-10.049 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	65.0 %HR	-10.560 %HR	± 5.0 %HR	± 1.9E+00 %HR