

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.865 °C	30.6 °C	-0.265 °C	± 1.6 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	40.0 %HR	9.839 %HR	± 1.6 %HR	± 1.9E+00 %HR
50 %HR	50.652 %HR	73.0 %HR	22.348 %HR	± 1.6 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	84.0 %HR	13.552 %HR	± 5.0 %HR	± 1.9E+00 %HR