

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.5033 µl	100 µl	0.4967 µl	± 0.5 µl	±	5.3E-02 µl
500.1222 µl	500 µl	-0.1222 µl	± 0.5 µl	±	5.7E-02 µl
999.6431 µl	1000 µl	0.3569 µl	± 0.5 µl	±	7.0E-02 µl