

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.06 RPM	100 RPM	-0.06 RPM	± 1 RPM	±	2.5E-01 RPM
499.58 RPM	500 RPM	0.42 RPM	± 1 RPM	±	6.1E-01 RPM
999.2 RPM	1000 RPM	0.8 RPM	± 1 RPM	±	1.6E+00 RPM
1499 RPM	1500 RPM	1 RPM	± 1 RPM	±	1.7E+00 RPM