

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.343 °C	30.5 °C	0.157 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.9 %HR	2.841 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	51.8 %HR	2.794 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.223 %HR	72.8 %HR	-2.423 %HR	± 3.0 %HR	± 1.9E+00 %HR