

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.343 °C	30.3 °C	-0.043 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.8 %HR	2.741 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	51.4 %HR	2.394 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.223 %HR	74.3 %HR	-0.923 %HR	± 3.0 %HR	± 1.9E+00 %HR