

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.333 °C	30.3 °C	-0.033 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.282 %HR	32.2 %HR	2.918 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.302 %HR	52.3 %HR	2.998 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	73.495 %HR	74.8 %HR	1.305 %HR	± 3.0 %HR	± 1.9E+00 %HR