

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.463 °C	30.1 °C	-0.363 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.314 %HR	32.3 %HR	1.986 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	52.0 %HR	1.818 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.3 %HR	-2.148 %HR	± 3.0 %HR	± 1.9E+00 %HR