

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.463 °C	30.0 °C	-0.463 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.314 %HR	32.9 %HR	2.586 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	51.0 %HR	0.818 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.2 %HR	-1.248 %HR	± 3.0 %HR	± 1.9E+00 %HR