

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.484 °C         | 30.3 °C   | -0.184 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.642 %HR        | 32.5 %HR  | 1.858 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.396 %HR        | 51.7 %HR  | 1.304 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 70.0 %HR  | -0.448 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |