

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.484 °C	30.1 °C	-0.384 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.642 %HR	28.7 %HR	-1.942 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	49.4 %HR	-0.996 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.5 %HR	-0.948 %HR	± 3.0 %HR	± 1.9E+00 %HR