

Patrón <i>(Standard)</i>		UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
7 Hz	0.283 g-rms	0.290 g-rms	0.007 g-rms	± 0.0085 g-rms	±	8.8E-03 g-rms
10 Hz	0.566 g-rms	0.564 g-rms	-0.002 g-rms	± 0.0170 g-rms	±	1.3E-02 g-rms
50 Hz	0.707 g-rms	0.708 g-rms	0.001 g-rms	± 0.0212 g-rms	±	1.5E-02 g-rms
100 Hz	0.707 g-rms	0.708 g-rms	0.001 g-rms	± 0.0212 g-rms	±	1.6E-02 g-rms
1000 Hz	0.707 g-rms	0.708 g-rms	0.001 g-rms	± 0.0212 g-rms	±	1.6E-02 g-rms
5000 Hz	0.707 g-rms	0.708 g-rms	0.001 g-rms	± 0.0212 g-rms	±	2.3E-02 g-rms