

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
2.5 RPM	2.5 RPM	0 RPM	± 0.10 RPM	±	5.8E-02 RPM
35 RPM	35.0 RPM	0 RPM	± 0.11 RPM	±	5.8E-02 RPM
150 RPM	149.9 RPM	-0.1 RPM	± 0.13 RPM	±	5.8E-02 RPM
580 RPM	579.9 RPM	-0.1 RPM	± 0.22 RPM	±	5.8E-02 RPM
800 RPM	799.9 RPM	-0.1 RPM	± 0.26 RPM	±	5.8E-01 RPM
1100 RPM	1099.9 RPM	-0.1 RPM	± 0.32 RPM	±	5.8E-02 RPM
99999 RPM	99997 RPM	-2 RPM	± 21 RPM	±	8.4E-01 RPM