

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-24.999 °C	-24.9 °C	0.099 °C	± 0.5 °C	± 2.2E-01 °C
0 °C	0.250 °C	0.6 °C	0.350 °C	± 0.5 °C	± 2.4E-01 °C
25 °C	25.249 °C	25.5 °C	0.251 °C	± 0.5 °C	± 1.7E-01 °C
50 °C	50.278 °C	50.7 °C	0.422 °C	± 0.5 °C	± 1.6E-01 °C
80 °C	80.331 °C	80.5 °C	0.169 °C	± 0.5 °C	± 1.6E-01 °C
30 %HR	30.263 %HR	31.3 %HR	1.037 %HR	± 3.0 %HR	± 1.0E+00 %HR
40 %HR	40.284 %HR	42.1 %HR	1.816 %HR	± 3.0 %HR	± 1.0E+00 %HR
50 %HR	50.335 %HR	50.6 %HR	0.265 %HR	± 3.0 %HR	± 1.2E+00 %HR
70 %HR	70.141 %HR	71.1 %HR	0.959 %HR	± 3.0 %HR	± 1.2E+00 %HR
80 %HR	80.264 %HR	81.3 %HR	1.036 %HR	± 3.0 %HR	± 1.1E+00 %HR