

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
23 °C	23.514 °C	23.8 °C	0.286 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.251 %HR	29.8 %HR	0.549 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.605 %HR	53.0 %HR	0.395 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.899 %HR	77.8 %HR	0.901 %HR	± 3.0 %HR	± 1.9E+00 %HR