

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
23 °C	23.514 °C	23.9 °C	0.386 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.251 %HR	30.0 %HR	0.749 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.605 %HR	53.4 %HR	0.795 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.899 %HR	78.0 %HR	1.101 %HR	± 3.0 %HR	± 1.9E+00 %HR