

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
23 °C	23.514 °C	23.9 °C	0.386 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.251 %HR	29.7 %HR	0.449 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	52.605 %HR	52.9 %HR	0.295 %HR	± 3 %HR	± 1.9E+00 %HR
75 %HR	76.899 %HR	77.3 %HR	0.401 %HR	± 3 %HR	± 1.9E+00 %HR