

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.297	0.29921	-2.210	± 1.4	± 4.3E-01 mm					
0 a 0.498	0.49922	-1.220	± 1.5	± 4.3E-01 mm					
0 a 0.999	0.99925	-0.250	± 1.9	± 4.3E-01 mm					
0 a 1.999	1.99929	-0.290	± 2.7	± 4.8E-01 mm					
0 a 5	4.99943	0.570	± 5.2	± 4.8E-01 mm					
0 a 10.001	9.99965	1.350	± 9.2	± 3.6E-01 mm					
0 a 15.001	14.99986	1.140	± 13.2	± 3.6E-01 mm					
0 a 20.002	20.00006	1.940	± 17.2	± 5.5E-01 mm					
0 a 30.002	30.00043	1.570	± 25.2	± 5.5E-01 mm					
0 a 50.001	50.00107	-0.070	± 41.2	± 6.3E-01 mm					