

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.3	0.297	3.790	± 1.4	± 4.3E-01 mm					
0 a 0.5	0.498	2.780	± 1.6	± 4.3E-01 mm					
0 a 1	0.999	1.750	± 2	± 4.3E-01 mm					
0 a 2	1.999	1.710	± 2.8	± 4.8E-01 mm					
0 a 5	5.000	0.570	± 5.2	± 4.8E-01 mm					
0 a 10	10.001	-0.650	± 9.2	± 3.6E-01 mm					
0 a 15	15.001	-0.860	± 13.2	± 3.6E-01 mm					
0 a 20	20.002	-2.060	± 17.2	± 3.6E-01 mm					
0 a 30	30.002	-2.430	± 25.2	± 5.5E-01 mm					
0 a 50	50.003	-4.070	± 41.2	± 6.3E-01 mm					