

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.751 °C	29.4 °C	-0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.648 %HR	30.6 %HR	1.952 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.977 %HR	50.2 %HR	0.223 %HR	± 2.5 %HR	± 1.9E+00 %HR
75 %HR	71.685 %HR	70.9 %HR	-0.785 %HR	± 2.5 %HR	± 1.9E+00 %HR