

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3599.955 s	3599.91 s	-0.045 s	-0.001%	± 0.0012 %	±	5.6E-02 s
3599.425 s	3599.39 s	-0.035 s	-0.001%	± 0.0012 %	±	5.6E-02 s
3599.218 s	3599.18 s	-0.038 s	-0.001%	± 0.0012 %	±	5.6E-02 s