

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.951 °C	15.1 °C	0.149 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.2 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.5 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.448 %HR	69.5 %HR	-0.948 %HR	± 3.0 %HR	± 1.9E+00 %HR