

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.406 °C	31.7 °C	0.294 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.969 %HR	31.5 %HR	0.531 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	52.1 %HR	2.542 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.558 %HR	76.7 %HR	2.142 %HR	± 5.0 %HR	± 1.9E+00 %HR