

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.406 °C	31.8 °C	0.394 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.969 %HR	32.0 %HR	1.031 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	53.5 %HR	3.942 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.558 %HR	75.8 %HR	1.242 %HR	± 5.0 %HR	± 1.9E+00 %HR