

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
23 °C	23.514 °C	24.2 °C	0.686 °C	±	1 °C	± 4.0E-01 °C
30 %HR	29.251 %HR	33 %HR	3.749 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	52.605 %HR	56 %HR	3.395 %HR	±	5 %HR	± 1.9E+00 %HR
75 %HR	76.899 %HR	80 %HR	3.101 %HR	±	5 %HR	± 1.9E+00 %HR