

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>		Incertidumbre <i>(Uncertainty)</i>
23 °C	23.514 °C	24.0 °C	0.486 °C	±	1 °C	± 4.0E-01 °C
30 %HR	29.251 %HR	34 %HR	4.749 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	52.605 %HR	57 %HR	4.395 %HR	±	5 %HR	± 1.9E+00 %HR
75 %HR	76.899 %HR	81 %HR	4.101 %HR	±	5 %HR	± 1.9E+00 %HR