

| Nominal <i>(Nominal)</i> | Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> |       | Incertidumbre <i>(Uncertainty)</i> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|-------|------------------------------------|
| 23 °C                    | 23.514 °C                | 24.3 °C          | 0.786 °C             | ±                           | 1 °C  | ± 4.0E-01 °C                       |
| 30 %HR                   | 29.251 %HR               | 34 %HR           | 4.749 %HR            | ±                           | 5 %HR | ± 1.9E+00 %HR                      |
| 50 %HR                   | 52.605 %HR               | 57 %HR           | 4.395 %HR            | ±                           | 5 %HR | ± 1.9E+00 %HR                      |
| 75 %HR                   | 76.899 %HR               | 81 %HR           | 4.101 %HR            | ±                           | 5 %HR | ± 1.9E+00 %HR                      |