

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.156 °C	2.2 °C	0.044 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	4.171 °C	4.2 °C	0.029 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	6.147 °C	6.2 °C	0.053 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.032 °C	8.1 °C	0.068 °C	± 0.5 °C	± 4.0E-01 °C