

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.256 °C	3.1 °C	0.844 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.588 °C	36.0 °C	0.412 °C	± 1.0 °C	± 4.0E-01 °C
# 15 %HR	14.302 %HR	15.8 %HR	1.498 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.726 %HR	68.8 %HR	-2.926 %HR	± 5.0 %HR	± 1.9E+00 %HR