

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.256 °C	3.0 °C	0.744 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.588 °C	36.0 °C	0.412 °C	± 1.0 °C	± 4.0E-01 °C
# 15 %HR	14.302 %HR	15.1 %HR	0.798 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.726 %HR	71.1 %HR	-0.626 %HR	± 5.0 %HR	± 1.9E+00 %HR