

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.751 °C	29.5 °C	-0.251 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	33.464 %HR	33 %HR	-0.464 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.977 %HR	50 %HR	0.023 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.685 %HR	75 %HR	3.315 %HR	± 5.0 %HR	± 1.9E+00 %HR