

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.672 %HR	27.8 %HR	-2.872 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.018 %HR	47.3 %HR	-2.718 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.396 %HR	70.9 %HR	0.504 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.284 %HR	92.0 %HR	1.716 %HR	± 5 %HR	± 1.9E+00 %HR

