

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-24.7 °C	0.339 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.341 °C	-0.6 °C	-0.259 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	4.7 °C	-0.444 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.977 °C	10.3 °C	0.323 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.720 °C	25.9 °C	0.180 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.834 °C	30.9 °C	0.066 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.722 °C	40.3 °C	-0.422 °C	± 0.5 °C	± 4.0E-01 °C

