

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-24.9 °C	0.139 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.432 °C	-0.4 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.033 °C	5.1 °C	0.067 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.957 °C	10.1 °C	0.143 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.868 °C	25.3 °C	0.432 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.291 °C	40.2 °C	-0.091 °C	± 0.5 °C	± 4.0E-01 °C

