

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.039 °C        | -25.4 °C  | -0.361 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.432 °C         | -0.6 °C   | -0.168 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.033 °C          | 4.9 °C    | -0.133 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.957 °C          | 9.9 °C    | -0.057 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.868 °C         | 25.0 °C   | 0.132 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 29.9 °C   | -0.162 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.291 °C         | 40.0 °C   | -0.291 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

