

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.039 °C	-25.2 °C	-0.161 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.341 °C	-0.4 °C	-0.059 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.144 °C	4.8 °C	-0.344 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.977 °C	9.8 °C	-0.177 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.720 °C	26.0 °C	0.280 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.834 °C	30.4 °C	-0.434 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.722 °C	40.3 °C	-0.422 °C	± 0.5 °C	± 4.0E-01 °C

