

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 20 °C                      | 20.105 °C                  | 20.2 °C            | 0.095 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 29.751 °C                  | 30.0 °C            | 0.249 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 50 %HR                     | 49.977 %HR                 | 51.8 %HR           | 1.823 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 71.685 %HR                 | 68.7 %HR           | -2.985 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |