

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 20 °C                    | 20.105 °C                | 20.3 °C          | 0.195 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 30 °C                    | 29.751 °C                | 30.0 °C          | 0.249 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 50 %HR                   | 49.977 %HR               | 51.9 %HR         | 1.923 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |
| 75 %HR                   | 71.685 %HR               | 71.4 %HR         | -0.285 %HR           | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |