

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.105 °C	20.2 °C	0.095 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.751 °C	30.1 °C	0.349 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.977 %HR	50.5 %HR	0.523 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	71.685 %HR	69.7 %HR	-1.985 %HR	± 3.0 %HR	± 1.9E+00 %HR