

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
cm	cm	in	in	(Uncertainty)		cm	cm	in	in	(Uncertainty)	
0 a 0.05	0.04988	0.0014	± 0.02	±	4.3E-04 in						
0 a 0.15	0.14989	0.0013	± 0.2	±	4.3E-04 in						
0 a 0.2	0.19988	0.0014	± 0.2	±	4.3E-04 in						
0 a 0.25	0.24992	0.0010	± 0.2	±	4.3E-04 in						
0 a 0.3	0.29989	0.0013	± 0.2	±	4.3E-04 in						
0 a 0.35	0.34989	0.0013	± 0.2	±	4.3E-04 in						
0 a 0.4	0.39990	0.0012	± 0.2	±	4.3E-04 in						
0 a 0.45	0.44990	0.0012	± 0.2	±	4.3E-04 in						
0 a 0.5	0.49992	0.0010	± 0.3	±	4.3E-04 in						
0 a 0.6	0.59992	0.0010	± 0.3	±	4.3E-04 in						