

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 0.1001 ml                | 0.1 ml           | -0.0001 ml           | ± 0.020 ml                  | ±                                  | 5.1E-02 µl |
| 1.0013 ml                | 1 ml             | -0.0013 ml           | ± 0.020 ml                  | ±                                  | 6.0E-02 µl |
| 5.0105 ml                | 5 ml             | -0.0105 ml           | ± 0.020 ml                  | ±                                  | 1.8E-01 µl |
| 10.0170 ml               | 10 ml            | -0.0170 ml           | ± 0.020 ml                  | ±                                  | 3.3E-01 µl |