

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.1001 ml	0.1 ml	-0.0001 ml	± 0.020 ml	±	5.1E-02 µl
1.0013 ml	1.0 ml	-0.0013 ml	± 0.020 ml	±	6.0E-02 µl
5.0105 ml	5.0 ml	-0.0105 ml	± 0.020 ml	±	1.8E-01 µl
10.0170 ml	10.0 ml	-0.0170 ml	± 0.020 ml	±	3.3E-01 µl