
| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (M.P.E) | Incertidumbre (Uncertainty) | |
|--------------------------|------------------|----------------------|----------------------|------------------------------------|------------|
| 1.0019 ml | 1 ml | -0.00189 ml | ± 0.006 ml | ± | 7.1E-02 µl |
| 2.0040 ml | 2 ml | -0.00398 ml | ± 0.010 ml | ± | 1.2E-01 µl |
| 4.9903 ml | 5 ml | 0.00974 ml | ± 0.015 ml | ± | 2.6E-01 µl |