

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0019 ml	1 ml	-0.00189 ml	± 0.006 ml	±	#REF! #REF!
2.0040 ml	2 ml	-0.00398 ml	± 0.010 ml	±	#REF! #REF!
4.9903 ml	5 ml	0.00974 ml	± 0.015 ml	±	#REF! #REF!